

Certificate of Analysis

Product Information

Part Number: 00654-00,
Description: Buffer 4.01 pH Red 500 mL
Lot Number: CC719389
Expiration Date: 06 Apr 2023
Method Of Analysis: pH

Actual Analysis

| <u>Acceptance Range</u> | <u>Results</u> | <u>Unit Of Measure</u> |
|----------------------------|----------------|------------------------|
| 4.01 pH +/-0.01 pH AT 25°C | 4.01 | pH AT 25°C |

Traceability

| <u>Description</u> | <u>Serial Number</u> | <u>Due Date</u> | <u>NIST Traceable Reference</u> |
|-----------------------------|----------------------|-----------------|---------------------------------|
| pH/Ion Meter | 658R067-N023 | | |
| pH Reference Material 4.005 | E1115820 | 23 Nov 2023 | 158-20 |
| Digital Thermometer | 102008443 | 22 Feb 2022 | 4000-11972359 |
| pH Electrode | 15293-F18 | | |

This certificate indicates Traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

Certified By: Cole-Parmer Instrument Company
Quality Assurance

Certificate Date: 06 Apr 2021