
Certificate of Analysis

Product Information

Part Number: 00654-04,
Description: Buffer 7.00 pH Yellow 500 mL
Lot Number: CC546825
Expiration Date: 27 Mar 2020
Method Of Analysis: pH

Actual Analysis

<u>Acceptance Range</u>	<u>Results</u>	<u>Unit Of Measure</u>
7.00 pH +/-0.01 pH AT 25°C	7.00	pH AT 25°C

Traceability

<u>Description</u>	<u>Serial Number</u>	<u>Due Date</u>	<u>NIST Traceable Reference</u>
pH/Ion Meter	658R067-N023		
pH Reference Material 7.0000	E1204116	03 May 2019	041-16
Digital Thermometer	102008443	22 Feb 2019	4000-9250399
pH Electrode	15293-F18		

This certificate indicates Traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

Certified By: Cole-Parmer Instrument Company
Quality Assurance

Certificate Date: 27 Mar 2018