


Cole-Parmer®

SC-200 Series

Sample Concentrator



Instruction Manual
TECH0002 Version 1.1

Cole-Parmer®
essentials

Contents

Introduction	3
General Description	3
Important Safety Advice	4
Installation	5
Operation	6
Preparation	6
Processing of Samples	8
Maintenance	8
Cleaning and Care	8
Carriage Friction	8
Replacing Sealing Pad	9
In Case of Contamination	9
Service and Repair	10
Warranty	10
Replacement Parts and Accessories	10
Diagram of Gas Chamber	11
Technical Specification	12
Customer and Technical Support	12
Solvent Evaporation Temperatures	13

Introduction

Thank you for purchasing this Cole-Parmer product. To get the best performance from the equipment, and for your own safety, please read these instructions carefully before use.

Before discarding the packaging check that all parts are present and correct.

When unpacking the unit, check that the following have been removed from the packing:

- ❖ Base board
- ❖ Column and carriage
- ❖ Gas chamber
- ❖ Instruction manual
- ❖ Warranty card

General Description

The Cole-Parmer Sample Concentrator has been designed to provide a quick and convenient method of accelerating the concentration by evaporation of solvents from samples prior to analysis. It is designed to work in conjunction with Cole-Parmer DB100/3, DB200/3, DB-3, and DB-3A Block Heaters*.

The Sample Concentrator consists of a gas chamber mounted above a Block Heater into which vessels containing samples are placed. Hyperdermic needles carry the gas down from the chamber into the test tubes. The samples are heated from below by the heater and the flow of gas directed over the surface of the samples displaces the evaporated solvent from above the liquid's surface. This significantly increases the rate of sample concentration.

The unit provides accurate and reproducible control over the positioning of the needles. The gas chamber can be removed from its stand to provide easy access to the samples, and it can be replaced quickly. The whole unit is compact enough for convenient use in a fume cupboard.

* The Block Heater is not provided and must be purchased separately.

Important Safety Advice

Users should be aware of the following safety advice:

- ❖ **DO NOT** use combustible substances near hot objects.
- ❖ **DO NOT** use the equipment in hazardous atmospheres.
- ❖ **HIGH TEMPERATURES ARE DANGEROUS** as they can cause serious burns to operators and ignite combustible material. Users should be aware of the following safety advice:
- ❖ **USE CARE AND WEAR PROTECTIVE GLOVES TO PROTECT HANDS.**
- ❖ **DO NOT** apply severe physical stress or use at temperatures above the rated value.


Symbols Defined



Installation

For packing purposes the Sample Concentrator is supplied unassembled in three parts; the base board, the column and carriage, and the gas chamber. The unit is re-assembled as follows:

1. Unscrew the column fastening knob from the column support bracket already fixed to the base board. Locate the column on the two pins in the support bracket. Screw the fastening knob through the rear of the column and tighten firmly.
2. Align the threaded hole on the back of the gas chamber with the mounting on the carriage. Tighten the locking knob to secure the gas chamber in position.
3. Connect the intake nozzle at the rear of the gas chamber to a suitable supply of gas via a pressure reducing valve. The gas pressure should be adjusted as required at the reducing valve but should not exceed 2 psi. Pressure in excess of 2 psi will cause gas to be vented and wasted.
4. Place the Cole-Parmer Block Heater onto the base of the Sample Concentrator. It is positioned by means of two spigots on the base board which align with two centrally located screw heads on the underside of the Dri-Block[®] unit.




Operation

Preparation

1. Place the aluminium insert blocks to be used into the Dri-Block[®] unit. The blocks should touch each other and be positioned centrally.
2. Loosen the locking knob on the carriage. Remove the gas chamber, disconnecting the gas pipe if necessary, and lay the chamber upside down on a suitable work surface.
3. Insert the hyperdermic needles in the pattern to suit the test tubes to be placed in the heater.


The needle guide holes are labelled with a letter corresponding with the type of insert block used in the heater. These are listed in the following table. As an example if you use the block for twelve 15mm test tubes (e.g. block number F3506) you would place a needle in all the holes labelled C.

Only Sample Concentrators with the part number SC-200-96 or a replacement head part number 6106467 will have the 96 well array.


<i>Tube size & diameter</i>	<i>Hole array</i>	<i>Hole label</i>
5 x 6 deep	30	A
4 x 5 deep	20	B
3 x 4 deep	12	C
8 hole off-set array, e.g. F4490 		D

Press each needle firmly, pointed end first, through the guide holes and through the sealing pad behind so that it is held firmly in position. Build up the desired pattern of needles to suit the arrangement of test tubes to be used. Do not put a needle in any position which will not have a corresponding sample to be concentrated as this will only waste gas. The needles may be removed and replaced many times as the pad is self sealing.


4. Replace the gas chamber onto the carriage and tighten the locking knob. Re-connect the gas pipe if necessary.



SC-200 matrix plate



SC-200-96 matrix plate



Inserting the Needles

Processing of Samples



Warning - remember high temperatures are dangerous. The heater, insert blocks, sample vessels, needles and samples may be very hot. Always take the precautions listed earlier in this manual.

1. By turning the carriage handle, on the right hand side, counter clockwise raise the gas chamber to take the needles well away from the heater.
2. Place the sample vessels in the insert blocks and lower the chamber so that the tips of the needles enter the test tubes. Position them at the required height above the liquid surface.
3. Set the heater to the required temperature and adjust the gas flow to the desired rate. Remember not to exceed 2 psi.
4. To increase the rate of concentration the chamber may be lowered as evaporation takes place so that the needles follow the level of the samples in the tubes. Using the scale on the column fine adjustments are easily made.
5. On completion switch off the gas flow and raise the chamber out of the way.
6. For operation of the Dri-Block[®] see the user's manual supplied separately.

Maintenance

Cleaning and care

With proper care and operation, the equipment should give reliable service, however contamination or general misuse may reduce the effective life of the product and could cause a hazard.

Preventative maintenance should include keeping the product clean by protecting it from spillage, contamination or corrosive environments. If in doubt, please confirm that any intended method of decontamination will not damage the equipment by contacting Antylia Scientific Ltd.

Your Sample Concentrator can be cleaned by wiping with a cloth dipped in soapy water. Ethanol and formaldehyde may also be used.

NOTE: Do not use solvents or abrasive cleaners for cleaning any parts of this equipment.

Carriage friction

Height adjustment is by means of a friction drive. To allow for wear the friction pressure may be increased by tightening the 4 screws on the rear of the carriage sufficiently to give adequate grip. Ensure that all four are tightened equally.

Replacing Sealing Pad

After many repeated needle insertions the self sealing property of the sealing pad may start to degrade. Spare pads can be obtained from your supplier or from Cole-Parmer. Replacing the pad is simply accomplished as follows:

- ❖ Remove the gas chamber from the carriage, disconnect the gas pipe and place upside down on a suitable work surface.
- ❖ Remove the four screws securing the matrix plate.
- ❖ Lift off the matrix plate.
- ❖ Lift off the sealing pad. Ensure that the support strip underneath the sealing pad remains in place.

- ❖ Replace the sealing pad with the new one. Reassembly is the reverse of the above procedure.

In Case of Contamination



WARNING: THE FOLLOWING PROCEDURE IS INTENDED AS A GUIDE. SHOULD SPILLAGE OF A TOXIC OR HAZARDOUS FLUID OCCUR, THEN ADDITIONAL SPECIAL PRECAUTIONS MAY BE NECESSARY.

If the equipment has been exposed to contamination, the Responsible Body is responsible for carrying out appropriate decontamination. If hazardous material has been spilt on or inside the equipment, decontamination should only be undertaken under the control of the Responsible Body with due recognition of possible hazards. Before using any cleaning or decontamination method, the Responsible Body should check with the manufacturer that the proposed method will not damage the equipment. Prior to further use, the Responsible Body shall check the electrical safety of the unit. Only if all safety requirements are met can the unit be used again.

NOTE: In the event of this equipment or any part of the unit becoming damaged or requiring service, the item(s) should be returned to the manufacturer for repair accompanied by a decontamination certificate. Copies of the Certificate are available from the Distributor/Manufacturer.

At the end of its service life, the product must be accompanied by a Decontamination Certificate.

Servicing and Repair

This product range does not require any routine servicing, instructions for cleaning and decontamination are included.

In the event of product failure, it is recommended that any repair is only undertaken by suitably qualified personnel. For advice, please contact Cole-Parmer quoting the model and serial number.

Only spare parts supplied by the manufacturer or its agent should be used. Fitting of non-approved parts may affect the performance of the safety features of the instrument.

If in doubt, please contact:

Antylia Scientific Ltd Ltd.
Beacon Road,
Stone, Staffordshire,
ST15 0SA, United Kingdom
Tel: +44 (0)1785 812121
Email: cpservice@antylia.com
Web: www.coleparmer.com

Warranty

Antylia Scientific Ltd Ltd. warrants this equipment to be free from defects in material and workmanship when used under normal laboratory conditions for the period specified on the enclosed warranty card. In the event of a justified claim, Antylia Scientific Ltd will replace any defective component or replace the unit free of charge.

This warranty does NOT apply if:

- ❖ Any repair has been made or attempted other than by the manufacturer or its agent.
- ❖ Any minor coating chips or scratches occur during normal use (i.e., wear and tear).
- ❖ Damage is caused by fire, accident, misuse, neglect, incorrect adjustment or repair, damage


caused by installation, adaptation, modification or fitting of non-approved parts.

Replacement Parts and Accessories

The following replacement parts and accessories may be obtained from your supplier or from Cole-Parmer.

Part Number	Description
F7209	Pack of 100 x 76mm long needles
F7210	Pack of 100 x 127mm long needles
FSC4NCS	Pack of PTFE coated needles, 100 x 76mm long
FSC4NCL	Pack of PTFE coated needles, 100 x 127mm long
6101608	Spigot
6101604	Sealing pad
6101605	Support strip
6106410	Standard, three block, matrix plate
6106467	96 well matrix plate
SC-200	Standard, three block, matrix
SC-200-96	96 well matrix

Diagram of Gas Chamber



Technical Specification

Maximum vertical travel	320 mm
Maximum gas pressure	2 psi
Maximum gas usage	15 litres/min
Gas	Any inert gas (often nitrogen)
Needle positions	Variable to suit Cole-Parmer Dri-Blocks®
Gas intake nozzle diameter	6.35 mm (¼")
Height	530 mm
Width	240 mm
Depth	295 mm
Weight	3.5 Kg

Customer and Technical Support

For help and support, contact:

Antylia Scientific Ltd Ltd.

Beacon Road,

Stone, Staffordshire,

ST15 0SA, United Kingdom

Tel: +44 (0)1785 812121

Customer Services: cpinfo@antylia.com

Sales: cpsales@coleparmer.com

Technical Support: cptechsupport@antylia.com

Warranty, Repairs and Service: cpervice@antylia.com

Web: www.coleparmer.com

Solvent Evaporation Temperatures

IUPAC Name	Common Name	Boiling point (°C) at 1013 mbar pressure	Pressure (mbar) at which b.p is 40 °C
2-Propanone	Acetone	56	556
1-Pentanol	Pentyl alcohol	137	11
Benzene	Benzene	80	236
1-Butanol	Butyl alcohol	118	25
2-Methyl-2-propanol	Tert-butyl alcohol	83	130
Chlorobenzene	Chlorobenzene	132	36
Chloroform	Trichloromethane	61	474
Cyclohexane	Cyclohexane	81	235
Ethoxyethane	Diethyl ether	35	Atmospheric
1,2-Dichloroethane	Ethylene chloride	83	210
1,2-Dichloroethene	Cis-acetylene dichloride	60	479
1,2-Dichloroethene	Trans-acetylene dichloride	48	751
2-Propan-2-yloxypropane	Isopropyl ether	68	375
1,4-Dioxane	Diethylene oxide	101	107
N,N-dimethylformamide	Dimethylformamide (DMF)	153	11
Acetic acid	Ethanoic acid	118	44
Ethanol	Alcohol	79	175
Ethyl acetate	Ethyl ester	77	40
Heptane	Dipropylmethane	98	120
Hexane	n-Hexane	69	335
Propan-2-ol	Isopropanol	82	137
3-Methyl-1-butanol	Isoamyl alcohol	130	14
Butan-2-one	Methylethylketone (MEK)	80	243
Methanol	Methyl alcohol	65	337
Dichloromethane	Methylene chloride	40	Atmospheric
Pentane	Pentane	36	Atmospheric
Propan-1-ol	Propyl alcohol	97	67
1,1,1,2,2-Pentachloroethane	Pentachloroethane	162	13
1,1,2,2-Tetrachloroethane	Tetrachloroethane	138	35
1,1,1-Trichloroethane	Trichloroethane	75	271
Tetrachloromethane	Carbon tetrachloride	76	300
1,1,2,2-Tetrachloroethene	Tetrachloroethylene	121	53
Oxolane	Tetrahydrofuran (THF)	67	357
Toluene	Methylbenzene	111	77
1,1,2-Trichloroethene	Trichloroethylene	87	183
Oxidane	Water	100	72
Dimethylbenzene	Xylene (mixed)	137	25
1,2-Dimethylbenzene	o-Xylene	144	34
1,3-Dimethylbenzene	m-Xylene	139	25
1,4-Dimethylbenzene	p-Xylene	138	31

NOTES

NOTES

NOTES



This product meets the applicable CE Directives and UKCA Legislation for radio frequency interference and may be expected not to interfere with, or be affected by, other equipment with similar qualifications. We cannot be sure that other equipment used in its vicinity will meet these standards and so we cannot guarantee

that interference will not occur in practise. Where there is a possibility that injury, damage or loss might occur if equipment malfunctions due to radio frequency interference, or for general advise before use, contact the manufacturer.

Declaration of Conformity is available to view online at www.coleparmer.com

EU Representative address

Antylia Scientific GmbH
Futtererstraße 16
97877 Wertheim
Deutschland
Tel: +49 9377 9203-0
Email: sales@coleparmer.de

UK Representative address

Antylia Scientific
9 Orion Court
Ambuscade Road
Colmworth Business Park
St. Neots
PE19 8YX
United Kingdom
Tel: +44 (0) 1480 277339
Email: enquiries@antylia.com

Ordering Information

Order No.	Series	Model	Legacy SKU
36620-40	SC-200	SC-200	FSC400D
36620-41	SC-200	SC-200-96	FSC496D

Warranty Registration



UK

T: +44 (0) 1480 272279
E: uk.sales@antylia.com
W: coleparmer.co.uk

India

T: +9122 61394444
E: info@coleparmer.in
W: coleparmer.in

Germany

T: +49 (0) 9377 92030
E: de.sales@antylia.com
W: coleparmer.de

China

T: +1 847 549 7600
E: sales@antylia.com
W: coleparmer.com

France

T: +33 (0) 1486 37800
E: fr.sales@antylia.com
W: coleparmer.fr

USA

T: +1 847 549 7600
E: sales@antylia.com
W: coleparmer.com

Italy

T: +39 (0) 284349215
E: it.sales@antylia.com
W: coleparmer.it

Canada

T: +514 355 6100
E: info@antylia.ca
W: coleparmer.ca

Other

T: +1 847 549 7600

Cole-Parmer®
essentials

Antylia Scientific Ltd.

Beacon Road,
Stone,
Staffordshire,
ST15 0SA,
United Kingdom

