

Cole-Parmer®

Product: Laboratory Equipment File Reference: TFR901071_001

Manufacturer: Cole Parmer Limited

Beacon Road

Stone

Staffordshire ST15 OSA United Kingdom

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration Automatic Melting Point Apparatus

MP-800D

The object of the declaration described above is in conformity with the relevant United Kingdom Legislation

Electromagnetic Compatibility Regulations 2016 No. 1091 Electrical Equipment (Safety) Regulations 2016 No. 1101

RoHS Regulations 2012 No. 3032 (including 2019 No. 492)

References to the relevant designated standards used or references to the other technical specifications in relation to which conformity is declared:

BS EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and

+A1:2019 laboratory use – Part 1: General requirements

BS EN 61010-2-010:2014 Safety requirements for electrical equipment for measurement, control and

laboratory use - Part 2-010: Particular requirements for laboratory equipment

for the heating of materials

BS EN 61326-1:2013 Electrical equipment for the measurement, control and Laboratory use – Part 1:

General requirements

Signed for and on behalf of the above manufacturer

Original Year of Marking 2022

Place of Issue Stone, UK

Authorised Representative Martin Card

Title EMEA Quality and Engineering Manager

Signature MM...

Date 10 OCTOBER 2022