

## **Certificate of Conformity**

**Part Number** 06286-32

**Description** Powder Sample Collector, Sterile, HDPE, 500mm length,

20mL; 20/PK

Lot Number 104713

Expiry Date 20 May 2027

Material HDPE (High Density Polyethylene)

Certificate Issued 14 September 2022

- We confirm that the goods are manufactured from HDPE that conforms to:
  - o FDA CFR 177.1520
  - EU Regulations 10/2011
  - o EC Regulations 1935/2004
- The HDPE resin used in the manufacture of the above products is BSE/TSE free. During manufacture this item
  does not come into contact with any products or processes that may pose a BSE/TSE risk.
- These goods are assembled and packed in a class 100,000 cleanroom. They are individually packed.
- This product is designed for single use only. It must not be used multiple times.
- Ultrasonic welding is used to adhere the components
- The information detailed above has been given to us by our suppliers, it is passed on in good faith and is to the best of our knowledge correct.

625 East Bunker Ct., Vernon Hills, IL 60061 Tel: +1.800.323.4340 or: +1.847.549.7600 coleparmer.com