

Certificate of Conformity

| Part Number | 06286-57 |
|--------------------|---|
| Description | Liquid Sample Collector, HDPE, 500mm length, 100mL; 20/PK |
| Lot Number | 104735 |
| Material | HDPE (High Density Polyethylene)Silicone Rubber O-ring |
| Certificate Issued | 14 September 2022 |

- We confirm that the goods are manufactured from HDPE that conforms to
 - FDA CFR 177.1520
 - EU Regulations 10/2011
 - EC Regulations 1935/2004
- The silicone rubber used for the O-ring seals conforms to FDA CRF 177.2600
- The HDPE resin used in the manufacture of the above products is BSE/TSE free. During manufacture this item does not come into contact with any products or processes that may pose a BSE/TSE risk.
- These goods are assembled and packed in a class 100,000 cleanroom. They are individually packed.
- This product is designed for single use only. It must not be used multiple times.
- The information detailed above has been given to us by our suppliers, it is passed on in good faith and is to the best of our knowledge correct.



Cole-Parmer 625 East Bunker Ct., Vernon Hills, IL 60061 Tel: +1.800.323.4340 or : +1.847.549.7600 coleparmer.com