

Part Number	06286-58
Description	Liquid Sample Collector, Sterile, HDPE, 500mm length, 100mL; 20/PK
Lot Number	104736
Expiry Date	09 May 2027
Material	<ul style="list-style-type: none">• HDPE (High Density Polyethylene)• Silicone Rubber O-ring
Certificate Issued	14 September 2022

- We confirm that the goods are manufactured from HDPE that conforms to
 - FDA CFR 177.1520
 - EU Regulations 10/2011
 - EC Regulations 1935/2004
- The silicone rubber used for the O-ring seals conforms to FDA CRF 177.2600
- The HDPE resin used in the manufacture of the above products is BSE/TSE free. During manufacture this item does not come into contact with any products or processes that may pose a BSE/TSE risk.
- These goods are assembled and packed in a class 100,000 cleanroom. They are individually packed.
- This product is designed for single use only. It must not be used multiple times.
- The information detailed above has been given to us by our suppliers, it is passed on in good faith and is to the best of our knowledge correct.