

## GAMMA IRRADIATION REPORT

**Product Name** : Media Bottle, PET, Sterile, 1000 mL  
**Volume** : 1000 ml  
**Catalogue Number** : 06287-03  
**Lot Number** : BPETS22090101  
**Radiation Date** : 26.08.2022 – 27.08.2022  
**Radiation Lot** : 220326  
**Sterilization Dose Required by Customer** : min. 10 kGy - max. 40 kGy

The product mentioned above had been irradiated by the <sup>60</sup>Co y-ray and received the following doses within the precision limit of the dosimetry system employed :

**Absorbed Dose** : 13,1 kGy  
**Dosimeter Lot No** : 201228