

Product: Laboratory Equipment

File Reference: TFR924121_002

Manufacturer: Cole-Parmer Limited
Beacon Road
Stone
Staffordshire ST15 0SA
United Kingdom

European Representative: Cole-Parmer Limited
STRADA PADANA SUPERIORE 290
20055 VIMODRONE (MILAN)
Italy

This declaration of conformity is issued under the sole responsibility of the manufacturer

**Object of Declaration Cole-Parmer Wrist-action flask shaker
WA-200 (51900-02)**


The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation

Electromagnetic Compatibility Directive 2014/30/EU
Low Voltage Directive 2014/35/EU
RoHS Directive 2011/65/EU (including CDD 2015/863)

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

BS EN 61010-1:2010 +A1:2019	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements
BS EN 61010-2-051:2020	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-051: Particular requirements for laboratory equipment for mixing and stirring
BS EN 61326-1:2021	Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements

Signed for and on behalf of the above manufacturer

Original Year of Marking	1996
Place of Issue	Stone, United Kingdom
Authorised Representative	Martin Card
Title	Senior Director, Global Engineering
Signature	
Date	23rd September 2025



Produit:	Équipement Laboratoire	Référence du Dossier:	TFR924121_002
Fabricant:	Cole-Parmer Limited Beacon Road Stone Staffordshire ST15 0SA United Kingdom	Représentant Européen:	Cole-Parmer Limited STRADA PADANA SUPERIORE 290 20055 VIMODRONE (MILAN) Italy

Cette déclaration de conformité est émise sous la responsabilité exclusive du fabricant

Objet de la Déclaration **Cole-Parmer Wrist-action flask shaker
WA-200 (51900-02)**

L'objet de la déclaration sus-mentionnée est conforme à la législation d'harmonisation de l'Union applicable

Electromagnetic Compatibility Directive	2014/30/EU
Low Voltage Directive	2014/35/EU
RoHS Directive	2011/65/EU (including CDD 2015/863)

Les Références aux normes harmonisées pertinentes utilisées ou les références aux autres spécifications techniques par rapport auxquelles la conformité est déclarée:

BS EN 61010-1:2010 +A1:2019	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements
--------------------------------	---

BS EN 61010-2-051:2020	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-051: Particular requirements for laboratory equipment for mixing and stirring
------------------------	---

BS EN 61326-1:2021	Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements
--------------------	---

Produkt: Laborausüstung Aktenzeichen: TFR924121_002

Hersteller: Cole-Parmer Limited
Beacon Road
Stone
Staffordshire ST15 0SA
United Kingdom

Europäischer Vertreter: Cole-Parmer Limited
STRADA PADANA SUPERIORE 290
20055 VIMODRONE (MILAN)
Italy

Diese Konformitätserklärung wird unter alleiniger Verantwortung des Herstellers ausgestellt


Gegenstand der Erklärung **Cole-Parmer Wrist-action flask shaker
WA-200 (51900-02)**

Der oben beschriebene Gegenstand der Erklärung entspricht der relevanten Unionsharmonisierungsgesetzgebung

Electromagnetic Compatibility Directive	2014/30/EU
Low Voltage Directive	2014/35/EU
RoHS Directive	2011/65/EU (einschließlich CDD 2015/863)

Verweise auf die relevanten harmonisierten Normen oder Verweise auf andere technische Spezifikationen, für die die Konformität erklärt wird:

- | | |
|--------------------------------|---|
| BS EN 61010-1:2010
+A1:2019 | Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements |
| BS EN 61010-2-051:2020 | Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-051: Particular requirements for laboratory equipment for mixing and stirring |
| BS EN 61326-1:2021 | Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements |



Prodotto: Attrezzatura da Laboratorio

Referenza: TFR924121_002

Produttore: Cole-Parmer Limited
Beacon Road
Stone
Staffordshire ST15 0SA
United Kingdom

Rappresentante Europeo: Cole-Parmer Limited
STRADA PADANA SUPERIORE 290
20055 VIMODRONE (MILAN)
Italy

La responsabilità della presente Dichiarazione di Conformità e' esclusivamente a carico del Produttore

**Oggetto della Dichiarazione Cole-Parmer Wrist-action flask shaker
WA-200 (51900-02)**

L'oggetto della dichiarazione sopra descritto è conforme alle Norme Armonizzate dell'Unione (Europea)

Electromagnetic Compatibility Directive	2014/30/EU
Low Voltage Directive	2014/35/EU
RoHS Directive	2011/65/EU (incluso CDD 2015/863)

In referenza alle pertinenti Norme Armonizzate e/o alle dichiarate Specifiche Tecniche conformi:

BS EN 61010-1:2010 +A1:2019	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements
--------------------------------	---

BS EN 61010-2-051:2020	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-051: Particular requirements for laboratory equipment for mixing and stirring
------------------------	---

BS EN 61326-1:2021	Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements
--------------------	---

Produto: Equipamento de Laboratório Referência de ficheiro: TFR924121_002

Fabricante: Cole-Parmer Limited Representante Cole-Parmer Limited
Beacon Road Europeu: STRADA PADANA SUPERIORE 290
Stone 20055 VIMODRONE (MILAN)
Staffordshire ST15 0SA Italy
United Kingdom

A presente declaração de conformidade é emitida sob a exclusiva responsabilidade do fabricante


**Objeto da declaração Cole-Parmer Wrist-action flask shaker
WA-200 (51900-02)**

O objeto da declaração acima descrita está em conformidade com a Legislação de Harmonização da União

Electromagnetic Compatibility Directive	2014/30/EU
Low Voltage Directive	2014/35/EU
RoHS Directive	2011/65/EU (incluindo CDD 2015/863)

Referências às normas harmonizadas aplicáveis utilizadas ou às outras especificações técnicas em relação às quais a conformidade é declarada:

BS EN 61010-1:2010 +A1:2019	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements
BS EN 61010-2-051:2020	Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-051: Particular requirements for laboratory equipment for mixing and stirring
BS EN 61326-1:2021	Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements



Producto:	Equipo de laboratorio	Referencia de archivo:	TFR924121_002
Fabricante:	Cole-Parmer Limited Beacon Road Stone Staffordshire ST15 0SA United Kingdom	Representante Europea:	Cole-Parmer Limited STRADA PADANA SUPERIORE 290 20055 VIMODRONE (MILAN) Italy

Esta declaración de conformidad se emite bajo la exclusiva del fabricante

Objeto de declaración **Cole-Parmer Wrist-action flask shaker
WA-200 (51900-02)**

El objeto de la declaración descrita anteriormente es conforme con la Legislación de Armonización de la Unión pertinente

Electromagnetic Compatibility Directive	2014/30/EU
Low Voltage Directive	2014/35/EU
RoHS Directive	2011/65/EU (including CDD 2015/863)

Referencias a las armonizadas pertinentes utilizadas o referencias a las demás especificaciones técnicas en relación con las cuales se declara la conformidad:

- | | |
|--------------------------------|---|
| BS EN 61010-1:2010
+A1:2019 | Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements |
| BS EN 61010-2-051:2020 | Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-051: Particular requirements for laboratory equipment for mixing and stirring |
| BS EN 61326-1:2021 | Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements |

