



Declaration of Conformity

Cole-Parmer®

Product: Laboratory Equipment

File Reference: TFR927091_001

Manufacturer: Cole Parmer Limited
Beacon Road
Stone
Staffordshire ST15 0SA
United Kingdom

This declaration of conformity is issued under the sole responsibility of the manufacturer

Objects of Declaration

**Cole-Parmer Tube Rotators
TR-200 series**

The objects of the declaration described above is in conformity with the relevant Union Harmonisation Legislation

Electromagnetic Compatibility Directive

2014/30/EU

Low Voltage Directive

2014/35/EU

RoHS Directive

2011/65/EU (including CDD 2015/863)

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

BS EN 61010-1:2010
+A1:2019

Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements

BS EN 61010-2-051:2015

Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-010: Particular requirements for laboratory equipment for mixing and stirring

BS EN 61326-1:2013

Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements

Signed for and on behalf of the above manufacturer

Original Year of Marking

2022

Place of Issue

Stone, UK

Authorised Representative

Martin Card

Title

EMEA Quality and Engineering Manager

Signature

Date

6 DEC 2022

Catalogue Code	Model	Description
51901-26	TR-200	Tube rotator, fixed speed, 230 V a.c.
51901-27	TR-200-120	Tube rotator, fixed speed, 120 V a.c.
51901-28	TR-200D	Tube rotator, variable speed, 230 V a.c.
51901-29	TR-200D-120	Tube rotator, variable speed, 120 V a.c.